

# DRINKING WATER TEST INSTRUMENTS

## Chlorometer 1000



The Chlorometer 1000 is a simple, direct reading, single parameter instrument for the determination of free and total chlorine using the Palintest DPD Method. The meter is compact and easy to use. Simply place the sample in the instrument and push the button. An accurate reading will instantly be available for your records. The instrument has a range of 0.00 - 5.00 mg/l Chlorine. The Chlorometer uses "AA" batteries and the keypad and screen are splash proof. The pocket-sized unit comes complete with a soft carrying case with belt loop.

PT 245/M2 - CHLOROMETER 1000 IN SOFT CASE .....	\$353.77
PT 233/M2 - CHLOROMETER CHECK STANDARDS .....	\$57.80

## Chlorometer Duo



The Chlorometer Duo is very similar to the Chlorometer, and lets water engineers and contractors measure the high chlorine levels used for pipework sterilisation as well as the low residual disinfectant levels in treated drinking water. Using the USEPA approved DPD Standard method, the unit has an operating range of 0.00 - 5.00 mg/l low range , 0 - 200 mg/l high range. It can also be used with the Palintest high range Chlorine HR test.

PT 227 - CHLOROMETER DUO IN SOFT CASE .....	\$419.72
PT 233/M2 - CHLOROMETER CHECK STANDARDS .....	\$57.80

## ChloroSense



ChloroSense uses a unique and patented electrochemical sensor to measure free and total chlorine simultaneously in a potable water sample. Taking less than a minute to give both results, the ChloroSense provides fast, accurate and portable chlorine measurement over a wide range of concentrations. The design of the new ChloroSense instrument combines rugged good looks with usability. Rated to IP67, the waterproof instrument will withstand the tough environments common in the water industry. It is self-contained, and no additional reagents or tubes are needed. Once the day's testing is over, download the results to a computer using the easy USB connection on the instrument. Chlorine monitoring has never been easier!

CS 100 - CHLOROSENSE .....	\$979.32
CS 110 - CHLOROSENSE ELECTRODES (100 TESTS) .....	\$101.74

All instruments are supplied with a test certificate confirming calibration. Rev. A 02-2009